

WHAT ABOUT A MAGNET?

- Asynchronous ?
- Turns off AICD ?
- No industry standard
- Usually does...
- Would not depend on it if you have time
 - Interrogate
 - Interrogate
 - Interrogate

Postoperative Management

(Anesthesiology 2011; 114:247-61)

- Interrogate and restore defibrillation function
- Observational study + case report
- Postoperative pacemaker check revealed the need to alter pacing mode or other parameters which include increasing ventricular thresholds
- **B2-B3 evidence**
- **My opinion**
 - Don't take the pads off until the device is interrogated and activated!

Conclusion

- Find out all the information you can
- Interrogate, interrogate, interrogate
- Prepare for the worst
 - These patients are sick
- Have a back up plan
- GO RAIDERS!